

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, October, 02

Composite spectrum for NIT2PRO_5T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 3.668×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: 2-Nitropropane, 2-NP, isonitro propane, nitroisopropane, sec-nitropropane, dimethylnitromethane, nipar s-30, nipar s-20, $\text{CH}_3\text{CHNO}_2\text{CH}_3$: [79-46-9]
- Physical properties: fw=89.094 g/mole, fp=-93° C, bp=120° C
- Supplier and stated purity: Aldrich, 97%
- Sample class: I (PNNL scale).
- Temperature of sample: 4.97 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760 ± 5 Torr
- Individual samples at 1.06100, 2.1938, 4.0655, 3.0530, 1.5532 and 0.73344 Torr. Path length = 19.94 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at -60 C to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 575 cm^{-1} (1.534 to 17.391 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\tau = 0.85$, $\nu = 530$)
- Composite spectrum created from 6 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.86%, Type B = 3%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.99999896 + 8.812 \times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)
- Baseline correction via 7th order polynomial subtraction