

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, July, 05

Composite spectrum for DECANE_25T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 5.8579×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: n-Decane, decane, decyl hydride, alkane C(10), $\text{CH}_3(\text{CH}_2)_8\text{CH}_3$: [124-18-5]
- Physical properties: MW=142.2838 g/mole, mp=-29.7° C, bp=174.4° C, Density (20 C) 0.730 g/cm³
- Supplier and stated purity: Aldrich, 99%
- Sample class: I (PNNL scale).
- Temperature of White cell (796.0 cm optical path length) 25 ± 2 C
- Diluent (high purity nitrogen) flowed at 24.2 liter/min (21.1° C), ambient atmospheric pressure 760 ± 5 Torr.
- Samples flowed at 5.000, 13.000, 3.000, 0.600, 9.000, 4.000, 22.000, 6.000, 16.000, 45.000, 1.000, 17.000, 24.000, 2.000 and 47.000 microliters/minute
- Individual samples at equivalent pressures of 0.019419, 0.050483, 0.011650, 0.002330, 0.034927, 0.015521, 0.085342, 0.023269, 0.062034, 0.174425, 0.003875, 0.065851, 0.092879, 0.007746 and 0.182009 Torr. Final data is a composite spectrum.
- Preparation: None

Instrument Parameters-

- Bruker-66V FTIR, evacuated optics bench.
- Modified to include second aperture, between interferometer output and White cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 580 cm^{-1} (1.538 to 17.241 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha=0.90$, $\epsilon=500$)
- Composite spectrum created from 15 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values ≥ 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.16%, Type B $\leq 7\%$

- Frequency correction (already applied): $V(\text{corrected})=V(\text{instrument})\cdot 0.99999959-3.45278\times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)
- Baseline correction via 7th order polynomial subtraction