

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : [sw.sharpe@pnl.gov](mailto:sw.sharpe@pnl.gov)

Version 1.0, August, 05

Composite spectrum for SECBUTBZ\_25T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum:  $5.5259 \times 10^{-6}$  grams/liter-meter

### Sample Conditions-

- Chemical name and CAS number: sec-Butylbenzene, 2-phenylbutane,  $C_6H_5C_4H_9$  : [138-98-8]
- Physical properties: MW=134.2206 g/mole, mp=-75° C, bp=173° C, Density (20 C) 0.862g/cm<sup>3</sup>
- Supplier and stated purity: Aldrich, 99+%
- Sample class: I (PNNL scale).
- Temperature of White cell (796.0 cm optical path length)  $25 \pm 2$  C
- Diluent (high purity nitrogen) flowed at 24.2 liter/min (21.1° C), ambient atmospheric pressure  $760 \pm 5$  Torr.
- Samples flowed at 25.000, 12.500, 7.000, 14.000, 10.000, 50.000, 37.500, 5.000, 2.000, 3.500, 4.000 and 1.000 microliters/minute
- Individual samples at equivalent pressures of 0.120915, 0.060449, 0.033852, 0.067676, 0.048347, 0.241669, 0.181228, 0.024157, 0.009662, 0.016908, 0.019316 and 0.004828 Torr.  
Final data is a composite spectrum.
- Preparation: None

### Instrument Parameters-

- Bruker-66V FTIR, evacuated optics bench.
- Modified to include second aperture, between interferometer output and White cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 600  $cm^{-1}$  (1.538 to 16.667 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112  $cm^{-1}$
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06  $cm^{-1}$
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0  $cm^{-1}$

### Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ( $\alpha=0.90$ ,  $\epsilon=500$ )
- Composite spectrum created from 12 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by  $T^2$  (transmission squared), all absorbance values  $\geq 1.6$  are given zero weight
- Calculated and estimated errors: Type A = 0.39%, Type B  $\leq 7\%$
- Frequency correction (already applied):  $V(\text{corrected})=V(\text{instrument}) * 0.99999959 - 3.45278 \times 10^{-4}$
- Axis units: X=wavenumbers ( $cm^{-1}$ ), Y=Absorbance (base-10)

- Baseline correction via 3<sup>rd</sup> order polynomial subtraction