

All data taken at Pacific Northwest National Laboratory (PNNL)
Operators: Timothy J. Johnson, Luisa T. Profeta and Robert L. Sams
Version 1.0, September 30, 2008

Composite spectrum for MVE_50T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 2.3912×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: Methyl vinyl ether; Ethene, methoxy-; Methoxyethylene; Methyl vinyl ether; Vinyl methyl ether; 1-Methoxyethylene; CH₂=CHOCH₃; Methoxyethene; C₃H₆O : [107-25-5]
- Physical properties: MW=58.0791 g/mole, mp= -121.6. °C, bp=5 °C, Density (25 C) = 0.744 g/cm³
- Supplier and stated purity: Customgas Solutions Inc., > 98 %
- Sample class: I (PNNL scale).
- Temperature of sample: 49.98 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760 ± 5 Torr.
- Individual samples at 1.03622, 3.7608, 4.0734, 28.60, 66.01 and 84.17 Torr. Path length= 19.96 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at 77K to remove air.

Instrument Parameters-

- Bruker-66V FTIR, evacuated optics bench.
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 550 cm⁻¹ (1.538 to 18.18 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm⁻¹
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm⁻¹
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm⁻¹

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha=0.90$, $\epsilon=500$)
- Composite spectrum created from 6 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T² (transmission squared), all absorbance values ≥ 1.6 are given zero weight
- Calculated and estimated errors: Type A =0.239%, Type B $\leq 3\%$
- Frequency correction (already applied): $V(\text{corrected})=V(\text{instrument})*1.00000135+8.5741 \times 10^{-4}$
- Axis units: X=wavenumbers (cm⁻¹), Y=Absorbance (base-10)

- Trace carbon dioxide and 0.5 % acetylene were removed by spectral subtraction and the composite spectrum was corrected by multiplying by 1.005059
- Baseline correction via 7th order polynomial subtraction