

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, May, 02

Composite spectrum for ME2BTE2_5T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 2.8875×10^{-6} grams/liter-meter

Unidentified contaminant peak observed at 891 cm^{-1} . Unable to correct or remove. See [ME2BTE2_5T.pdf](#) file.

Sample Conditions-

- Chemical name and CAS number: 2-Methyl-2-butene, n-amylene, -isoamylene, trimethylethylene, 2-methylbut-2-ene, 3-methyl-2-butene, $\text{H}_3\text{CCH}=\text{C}(\text{CH}_3)_2$: [513-35-9]
- Physical properties: fw=70.134 g/mole, fp=-134° C, bp=39° C
- Supplier and stated purity: Aldrich, 99%
- Sample class: I (PNNL scale).
- Temperature of sample: $5.05 \pm 0.02 \text{ C}$
- Diluent: Sample back filled with ultra high purity nitrogen to $760 \pm 5 \text{ Torr}$
- Individual samples at 1.01092, 4.4976, 16.62, 2.1397, 71.08, 33.38 and 8.7378 Torr. Path length = 19.94 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at 77 K to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: $6,500$ to 600 cm^{-1} (1.534 to 16.667 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\tau = 0.85$, $\nu = 530$)
- Composite spectrum created from 7 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.95%, Type B = 3%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.99999896 + 8.812 \times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)
- Baseline correction via 7th order polynomial subtraction