

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, July, 02

Composite spectrum for TCE1122_50T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 6.9105×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: 1,1,2,2-Tetrachloroethane, sym-tetrachloroethane, acetylene tetrachloride, dichloro 2,2-dichloroethane, Cellon, Bonoform, Westron, Acetsol, $\text{CHCl}_2\text{CHCl}_2$: [79-34-5]
- Physical properties: fw=167.8498 g/mole, fp=-43° C, bp=146.3° C
- Supplier and stated purity: Aldrich, 98%. ~1.33% tetrachloroethylene contamination observed
- Sample class: I (PNNL scale).
- Temperature of sample: 50.04 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760 ± 5 Torr
- Individual samples at 1.1440, 4.0474, 0.82222, 5.2121, 2.0545 and 3.0300 Torr. Path length = 19.96 cm. Final data is a composite spectrum.
- Preparation: Pump on liquid at -40 C to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: $6,500$ to 530 cm^{-1} (1.534 to 18.868 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha=0.90$, $\beta=500$)
- Composite spectrum created from 6 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.53%, Type B = 3%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.999998 - 2.75000e-6$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)
- Tetrachloroethylene features removed by spectral subtraction and composite spectrum rescaled to account for 1.33% contamination of tetrachloroethylene.

- Baseline correction via 7th order polynomial subtraction